Medicines & Healthcare products Regulatory Agency

MHRA

Medicines and Healthcare products Regulatory Agency

10 South Colonnade
Canary Wharf
London
E14 4PU
United Kingdom

+44 (0) 20 3080 6000

gov.uk/mhra

Tony Booley

Covestus Ltd
I Pulteney Road
Bath BA2 4HD
UK

3/4/20

Dear Tony,

[redacted]

From a regulatory point of view, we can confirm that the above [redacted] branded product can be placed on the UK market for professional use **only and not for home use**.

[redacted]

Legally the product can be placed on the market but just to confirm; MHRA **does not** quality approve the validity of the test.

It is not allowed to be available on-line.

If you are looking to supply this test kit to the health service contact covidtestingtriage@dhsc.gov.uk in the first instance. A one line of expression of interest will suffice. This then triggers an acknowledgement and a request to complete an excel form that is used for triage.

Yours sincerely

Rob Higgins
Senior Regulatory Affairs Manager
MHRA (Devices)
10 South Colonnade,
Canary Wharf,
London
E14 4PU

Tel : +44 (0) 20 3080 7185
Email : rob.higgins@mhra.gov.uk